



INTER UNIVERSITY CENTRE FOR BIOMEDICAL RESEARCH & SUPER SPECIALITY HOSPITAL



(An Autonomous Institution of Govt. of Kerala under Mahatma Gandhi University)
Rubber Board P. O., Thalappady, Kottayam, Kerala - 686 009, India

No. 43/A1/2017/IUCBR&SSH

Dated: 13.09.2017

SHORT QUOTATION NOTICE

Sealed quotations are invited for the supply of the following stores:

Sl. No	Name of Equipment	Specification	Nos
1	Autoclave	<ol style="list-style-type: none">1. Fully Automatic Vertical Top Loading Autoclave with capacity of 75 Litre, including automatic lid opening (pedal free)2. Chamber dimension: 350x780 mm minimum3. Usable Size: 350 x 680 mm4. Heater Power: 2.7 kW5. Temperature Range: 120 °C and above with Resolution 1.0 °C (User Settable)6. Operating Pressure: 15 psi or more7. Equipment should have the sterilizer timer, Drying Timer and safety interlock to prevent lid opening when a cycle is in program8. Digital displays for temperature & Time9. Digital Timer up to 99 minutes for drying10. All wetted parts should be made of Stainless steel including the Lid.11. Epoxy coating on steel exterior12. Low water detection and alarm, with auto cut off13. External bottle for draining water & the condensing steam.14. Microprocessor Based Control for ensuring repeatability & high accuracy.15. Spring loaded safety valve for over pressure.16. Independent self-resetting Thermal Safety Cut-out17. Pressure interlock on door.18. Equipment should be supplied with 2 Nos. of Dressing drum and one number of carrier with partition.	02 (Two)

1
0481-2353730



		19. Equipment should confirm CE Certification. 20. Three years warranty and additional 2 years Service agreement. 21. Power 220-240 V; 50 Hz (The equipment should have adapters for Indian type of electrical points.)	
--	--	---	--

The envelopes containing the quotation should bear the reference number and superscription **“Quotation for Autoclave to IUCBR&SSH”** and should be addressed to **Director, IUCBR&SSH, SME Campus Thalappady, Rubber Board P. O., Kottayam - 686009**. Intending Tenderers may submit the quotations on their own stationery.

Last date for receipt of quotations is **28.09.2017 at 02.00 PM**. Late quotations will not be accepted. The quotations will be opened at **02.30 PM on 28.09.2017** in the presence of the Tenderers or their authorized representatives who may be present at that time. The maximum period required for delivery of the articles should also be mentioned.

Details of the requirements and the conditions governing their supply can be obtained from Ph. No. 04812353730 till **28.09.2017 at 11.00 AM**.

Place: Thalappady
Date: 13.09.2017


13/9/17
Director, IUCBR&SSH


Copy to:

All Distributors and Suppliers of Small Lab Equipments in Kerala
Notice Board of MG University
Notice Board of IUCBR&SSH